

# WATO EX-65

## Anesthesia Workstation

### Physical Specifications

#### Dimensions and Weight

Height	1370 mm
Width	780 mm (not including breathing system)
	945 mm (including breathing system)
Depth	690 mm
Weight	<145 kg (without vaporizers and cylinders)

#### Top Shelf

Weight limit	30 kg
Width	305 mm
Length	545 mm

#### Work Surface

Height	850 mm
Area	1635 cm <sup>2</sup>

#### Drawer (3Xdrawers, Internal Dimension)

Height	130 mm
Width	415 mm
Depth	320 mm

#### Bag Arm

Height	1150 mm
Length	312 mm
Connection	ISO 22mm OD, 15mm ID

#### Casters

Diameter	125 mm
Brakes	Centre brake system with Lock / Unlock icons

### Ventilator Specifications

#### Modes of Ventilation

Manual/Spontaneous Ventilation/Bypass  
Volume Control Ventilation (VCV) with PLV function  
Pressure Control Ventilation (PCV) with/without volume guarantee (VG)  
Synchronized Intermittent Mandatory Ventilation (SIMV-Volume Controlled and SIMV-Pressure Controlled)  
Pressure Support Ventilation (PS) with apnea backup  
Synchronized Intermittent Mandatory Ventilation Volume Guarantee (SIMV-VG)  
Continuous Positive Airway Pressure/Pressure Support Ventilation (CPAP/PS)

#### Compensation

Circuit gas leakage compensation and automatic compliance compensation

#### Ventilation Parameters Range

Patient Size	Adult, Pediatric, Infant
Tidal volume	20~1500 mL (Volume Mode) (increments of 1 mL)
	5~1500 mL (Pressure Mode)
Pinsp	5~70 cmH <sub>2</sub> O (increments of 1 cmH <sub>2</sub> O)
Plimit	10~100 cmH <sub>2</sub> O (increments of 1 cmH <sub>2</sub> O)



$\Delta$ Psupp	3~60 cmH <sub>2</sub> O (increments of 1 cmH <sub>2</sub> O) 0, 3~60 cmH <sub>2</sub> O (CPAP/PS)
Rate	4~100 bpm (increments of 1 bpm)
I:E	4:1 - 1:8 (increments of 0.5)
Inspiratory pause (Tip:Ti)	OFF, 5% - 60% (increments of 1%)
Inspiratory time (T <sub>insp</sub> )	0.2 - 5.0 s (increments of 0.1 s)
Trigger window	5% - 90% (increments of 5%)
Flow trigger	0.5 ~ 15 L/min (increments of 0.5L/min)
Pressure trigger	-20~-1 cmH <sub>2</sub> O (increments of 1 cmH <sub>2</sub> O)
Expiration termination level	5% - 60% (increments of 5%)
Min Rate	2 - 60 bpm (increments of 1 bpm)
Tslope	0.0 - 2.0 s (increments of 0.1 s)
Apnea I: E	4:1~1:8 (increments of 0.5)
$\Delta$ Papnea	3 - 60 cmH <sub>2</sub> O (increments of 1 cmH <sub>2</sub> O)

#### Positive End Expiratory Pressure (PEEP)

Type	Integrated, electronic controlled
Range	OFF, 3~30 cmH <sub>2</sub> O (increments of 1 cm H <sub>2</sub> O)

#### Ventilator Performance

Driving pressure	280 kPa to 600 kPa
Peak gas flow	120 L/min + Fresh Gas Flow

#### Monitoring Parameters

Minute volume	0 ~ 100 L/min
Tidal volume	0~2500 ml
Inspired oxygen (FiO <sub>2</sub> )	18% ~ 100%
Peak airway pressure	-20 ~ 120 cmH <sub>2</sub> O
Mean pressure	-20 ~ 120 cmH <sub>2</sub> O
Plateau pressure	-20 ~ 120 cmH <sub>2</sub> O
I:E	4:1 ~ 1:10
Rate	0 ~ 120 bpm
PEEP	0 ~ 70 cmH <sub>2</sub> O
Resistance (R)	0 ~ 600 cmH <sub>2</sub> O/(L/s)
Compliance (C)	0 ~ 300 ml/cmH <sub>2</sub> O

#### Control Accuracy

Volume delivery	< 75 ml: $\pm$ 15 ml $\geq$ 75 ml: $\pm$ 15 ml or $\pm$ 10% of the set value, whichever is greater
Pinsp	$\pm$ 3.0 cmH <sub>2</sub> O or $\pm$ 8% of the set value, whichever is greater
Plimit	$\pm$ 3.0 cmH <sub>2</sub> O or $\pm$ 8% of the set value, whichever is greater
$\Delta$ Psupp	$\pm$ 3.0 cmH <sub>2</sub> O or $\pm$ 8% of the set value, whichever is greater
$\Delta$ Papnea	$\pm$ 3.0 cmH <sub>2</sub> O or $\pm$ 8% of the set value, whichever is greater
PEEP	OFF: $\leq$ 4.0 cmH <sub>2</sub> O 3 to 30 cmH <sub>2</sub> O: $\pm$ 2.0 cmH <sub>2</sub> O, or $\pm$ 8% of the set value, whichever is greater
Rate	$\pm$ 1bpm or $\pm$ 5% of the set value, whichever

I:E	is greater 2:1 to 1:4: $\pm 10\%$ of the set value Other range: $\pm 25\%$ of the set value
Tip:Ti	$\pm 8\%$
Tinsp	$\pm 0.2s$
<b>Monitoring Accuracy</b>	
Volume monitoring	< 75 ml: $\pm 15$ ml $\geq 75$ ml: $\pm 15$ ml or $\pm 10\%$ of the reading, whichever is greater
Pressure monitoring	$\pm 2.0$ cmH <sub>2</sub> O
PEEP	$\pm 2.0$ cmH <sub>2</sub> O or or $\pm 10\%$ of the reading, whichever is greater
Rate	$\pm 1$ bpm or $\pm 5\%$ of the reading, whichever is greater
I:E	2:1 to 1:4: $\pm 10\%$ of the reading Other range: $\pm 25\%$ of the reading
MV	$\pm 15\%$ of the reading

### Trend Graph

Continuous trend information with time discrete events for the latest 48 hours

### Trend Table

Continuous trend information together with time discrete events for the latest 48 hours

### Alarm Log Book

500 events storage, first in first out

### Alarm setting

Tidal volume	Low: 0 ~ 1595 ml High: 5 ~ 1600 ml
Minute volume	Low: 0 ~ 99 L/min High: 0.2 ~ 100 L/min
Inspired oxygen	Low: 18% ~ 98% High: OFF, 20% ~ 100%
Apnea alarm	VTe < 10ml measured in 20s Paw < (PEEP + 3) cmH <sub>2</sub> O in 20s
Airway pressure low	0 ~ 98 cmH <sub>2</sub> O
Airway pressure high	2 ~ 100 cmH <sub>2</sub> O
Sustained airway pressure alarm:	15s
Subatmospheric pressure alarm:	Paw < -10 cmH <sub>2</sub> O
Alarm silence countdown timer:	120 to 0 seconds

### Ventilator Components

#### Flow Sensor

Type	Variable orifice flow sensor
Location	Inspiratory and expiratory port

#### Oxygen Sensor

Type	Galvanic fuel cell
FiO <sub>2</sub> displayed	18% to 100%
Accuracy	$\pm$ (volume fraction of 2.5 % +2.5 % gas level)
Response Time	$\leq 20$ seconds

### Ventilator Screen

Display type	Color active matrix TFT touch screen
Display size	12.1 in diagonal
Pixel format	1024 x 768
Brightness	Adjustable
Screen display	configurable
Display parameters	All setting and alarm parameters (including Breath rate, I/E ratio, Tidal volume, Minute volume, PEEP, MEAN, PEAK, PLAT, and O <sub>2</sub> concentration, EtCO <sub>2</sub> , N <sub>2</sub> O, Aesthesia gas concentration, BIS)

Display waveforms	P-T, F-T, V-T, CO <sub>2</sub> , BIS, O <sub>2</sub> , Anesthetic gas, N <sub>2</sub> O
Spirometry loops	P-V, F-V and F-P
Timer	On screen timer

### Communication Ports

One RS-232C connector and one DB9 connector
Ethernet (RJ-45)
USB
VGA

### Vaporizers

Vaporizer	Mindray V60 Anesthetic Vaporizer or Penlon Sigma Delta Anesthetic Vaporizer
Support agents	Halothane, Enflurane, Isoflurane, Sevoflurane
Position	MAX.2
Mounting mode	Selectatec®, with interlocking function Plug-in®, with interlocking function

### Modules

#### Anesthesia Gas (AG) Module

Measurement mode	Infrared absorption
Monitor gases	CO <sub>2</sub> , N <sub>2</sub> O, Halothane, Enflurane, Isoflurane, Sevoflurane, Desflurane, MAC, Paramagnetic O <sub>2</sub> (optional)
Warm-up time	45 s (ISO accuracy mode) 10min (full accuracy mode)
Sample rate	Adu/Ped: 150, 180, 200 ml/min Neo: 100, 110, 120 ml/min
Accuracy	$\pm 10$ mL/min or $\pm 10\%$ of the set value, whichever is greater
Range	CO <sub>2</sub> : 0% ~ 10% Des: 0% ~ 18 % Sev: 0% ~ 8% Enf, Iso, Hal: 0% ~ 5% O <sub>2</sub> /N <sub>2</sub> O: 0% ~ 100%

#### Carbon Dioxide (CO<sub>2</sub>) Modules

Method	Infrared absorption
Module type	Mindray side-stream Capnostat mainstream Oridion micro-stream (optional)
Work mode	Standby or measurement
Displayed numerics	EtCO <sub>2</sub> , FiCO <sub>2</sub>
Waveform	Capnography

#### Side-Stream Carbon Dioxide (CO<sub>2</sub>) Module

Measurement range	0 ~ 99 mmHg
Accuracy	$\pm 2$ mmHg (0 ~ 40 mmHg) $\pm 5\%$ of the reading (41 ~ 76 mmHg) $\pm 10\%$ of the reading (77 ~ 99 mmHg)
Resolution	1 mmHg
Sampling rate	Neonatal: 100 mL/min and 120 mL/min optional Adult/children: 120 mL/min and 150 mL/min optional
Sampling rate accuracy:	$\pm 15\%$ of the set value or $\pm 15$ mL/min, whichever is greater
Warming-up time	< 1 min, enter the ISO accuracy mode After 1 min, enters the full accuracy mode
Response time	<4.5 s@100 mL/min <4.5 s@120 mL/min Measured by using neonatal watertrap and 2.5 m neonatal sampling line

<5.5 s@120 mL/min  
 <5 s@150 mL/min  
 Measured by using adult watertrap and 2.5 m adult sampling line

100 ~ 240V < 500 µA

### Capnostat Mainstream CO<sub>2</sub> Module

Measurement range 0 ~ 150 mmHg  
 Accuracy ± 2 mmHg (0 ~ 40 mmHg)  
 ± 5% of the reading (41 ~ 70 mmHg)  
 ± 8% of the reading (71 ~ 100 mmHg)  
 ± 10% of the reading (101 ~ 150 mmHg)  
 Resolution 1 mmHg  
 Rise time <60 ms  
 Response time <2 s  
 Alarm limit EtCO<sub>2</sub> High: OFF, 2 ~ 150 mmHg  
 EtCO<sub>2</sub> Low: OFF, 0 ~ 148 mmHg  
 FiCO<sub>2</sub> High: OFF, 1 ~ 150 mmHg

### Micro-stream CO<sub>2</sub> Module

Measurement range 0 ~ 99 mmHg  
 Accuracy 0 ~ 38 mmHg: ± 2 mmHg  
 39 ~ 99 mmHg: ± (5 % of the reading + 0.08 % of (the reading minus 38 mmHg))  
 Sampling rate 50 ml/min  
 Sampling accuracy -7.5 ml/min ~ + 15 ml/min  
 Initialization time 30s  
 Response time 2.9s  
 Rising time < 190 ms  
 Delay time 2.7s  
 Alarm range EtCO<sub>2</sub> High: OFF, 2 ~ 99 mmHg  
 EtCO<sub>2</sub> Low: OFF, 0 ~ 97 mmHg  
 FiCO<sub>2</sub> High: OFF, 1 ~ 99 mmHg

### BIS/BISx4 Module

Measured parameters EEG  
 BIS/BIS L, BIS R 0 ~ 100  
 Sweep speed 6.25 mm/s, 12.5 mm/s, 25 mm/s or 50 mm/s  
 Input impedance > 50 Mohm  
 Noise (RTI) < 0.3 µV (0.25 ~ 50 Hz)  
 Input signal range ± 1 mv  
 EEG bandwidth 0.25 ~ 100 Hz  
 Patient leakage current < 10 µA  
 Alarm limit BIS high: 2 ~ 100  
 BIS low: 0 ~ 98  
 Calculated parameters SQI/SQI L, SQI R; EMG/EMG L, EMG R; SR/SR L, SR R; SEF/SEF L, SEF R; TP/TP L, TP R; BC/BC L, BC R; sBIS L, sBIS R; sEMG L, sEMG R; ASYM  
 Impedance range 0 ~ 999 Kohm

### Agent Consumption Calculation

Calculation range 0 to 3000 ml  
 Accuracy ± 2 mL, or ± 15% of the reading, whichever is larger

### Agent consumption speed

Anesthetic agents Desflurane, Enflurane, Isoflurane, Sevoflurane and Halothane  
 Consumption speed Desflurane: 0 ~ 900 ml/h  
 Sevoflurane: 0 ~ 450 ml/h  
 Enflurane, Isoflurane and Halothane: 0 ~ 250 ml/h  
 Accuracy ± 2ml/h or ±15% of the displayed value, whichever is greater

### Electrical Specifications

#### Current Leakage

### Power And Battery Backup

Power input 220-240 Vac, 50/60 Hz, 6A  
 100-120 Vac, 50/60 Hz, 7A  
 100-240 Vac, 50/60 Hz, 7A  
 Auxiliary electrical outlets Up to 4 outlets (3A for each, total 5A)  
 Battery backup 90 minutes for 1 piece of battery (powered by new fully-charged batteries with 25°C ambient temperature)  
 240 minutes for 2 pieces of battery (powered by new fully-charged batteries with 25°C ambient temperature)  
 Battery type Build-in Li-ion battery, 11.1 VDC, 4500 mAh  
 Safety feature In case of electricity and battery failure, manual ventilation, gas delivery and agent delivery are possible

### Pneumatic Specifications

#### ACGO (Auxiliary Common Gas Outlet, Integrated)

Connector ISO 22 mm OD and 15 mm ID

#### Pipeline Supply

Gas type O<sub>2</sub>, N<sub>2</sub>O and Air  
 Pipeline input range 280 to 600 kPa  
 Pipeline connections DISS or NIST

#### Pipeline Supply Pressure Gauges

Display type Mechanical  
 Ranges 0 to 1000kPa  
 Accuracy ± (4% of the full scale reading + 8% of the actual reading)

#### Cylinder Supply

Cylinder Supply E Cylinder (American style or UK style)  
 O<sub>2</sub> Input Range 6.9 to 20 MPa  
 N<sub>2</sub>O Input Range 4.2 to 6 MPa  
 Air Input Range 6.9 to 20 MPa  
 Cylinder Connections Pin-Index Safety System (PISS)  
 Yoke Configuration O<sub>2</sub>, N<sub>2</sub>O, Air

#### Cylinder Supply Pressure Gauges

Display type Mechanical  
 Air Range 0 to 25 MPa  
 O<sub>2</sub> Range 0 to 25 MPa  
 N<sub>2</sub>O Range 0 to 10 MPa  
 Accuracy ± (4% of the full scale reading+8% of the actual reading)

#### O<sub>2</sub> Controls

Method N<sub>2</sub>O shut off with loss of O<sub>2</sub> pressure  
 Supply failure alarm ≤ 220.6 kPa  
 O<sub>2</sub> Flush 25 ~ 75 L/min

#### O<sub>2</sub>-N<sub>2</sub>O Link system

Type Mechanical  
 Range O<sub>2</sub> concentration not lower than 25%

#### Auxiliary O<sub>2</sub> Flowmeter

Range 0 ~ 15 L/min  
 Indicator Flow tube

#### Electronic Flow Meters

O<sub>2</sub> flow range 0 to 15 L/min

Air flow range	0 to 15 L/min
N <sub>2</sub> O flow range	0 to 10 L/min
Accuracy	between -10% and +10% of the indicated value (under 20°C and 101.3 kPa, for flow between 10% and 100% of full scale)

### Environmental Specifications

#### Operating

Temperature	10 ~ 40°C
Relative humidity	15% ~ 95% (noncondensing)
Barometric (Kpa)	70 ~ 106 kPa

#### Storage

Temperature	-20 ~ 60°C for main unit, -20 ~ 50°C for O <sub>2</sub> sensor
Relative humidity	10% ~ 95% (noncondensing)
Barometric	50 ~ 106 kPa

### Electromagnetic Compatibility

Immunity	Complies with all requirements of IEC 60601-1-2
Emissions	Complies with all requirements of IEC 60601-1-2

### Breathing System Specification

#### Breathing system volume (Pre-pak)

Automatic ventilation	2850 ml
Manual ventilation	1800 ml

#### Breathing system volume (Non Pre-pak)

Automatic ventilation	2600 ml
Manual ventilation	1800 ml

### System Components

Carbon dioxide absorbent canister	Absorbent capacity: 1500 mL
Integrated expiratory limb water trap	Capacity: 6 mL

### Breathing Circuit Parameters

Compliance	≤4 mL/100Pa (bag mode) Automatically compensates for compression losses within the breathing circuit in mechanical mode
Expiration resistance	< 6.0 cm H <sub>2</sub> O @60 L/min
Inspiration resistance	< 6.0 cm H <sub>2</sub> O @60 L/min

### System Pressure Gauge

Range	-20 ~ 100 cmH <sub>2</sub> O
Accuracy	± (2% of the full scale reading + 4% of the actual reading)

### Ports and Connectors

Exhalation	22 mm OD / 15 mm ID conical
Inhalation	22 mm OD /15 mm ID conical
Manual bag port	22 mm OD /15 mm ID conical

### Bag-to-Ventilator Switch

Type	Bi-stable
Control	Switch between manual and mechanical ventilation

### Integrated Adjustable Pressure Limiting (APL) Valve

Range	SP, 5 ~ 75 cmH <sub>2</sub> O
Tactile knob indication at above 30 cmH <sub>2</sub> O	
Accuracy	± 10 cmH <sub>2</sub> O or ± 15% of the setting value, which is greater

### Anesthetic Gas Scavenging System (AGSS)

Size (H x W x D)	430 x 132 x 114 mm
Type of disposal system	Active: High-flow or Low-flow Passive
Applicable standard	ISO 80601-2-13
Pump rate	75 ~ 105 L/min (High-flow) 25 ~ 50 L/min (Low-flow)

Pressure relief device: Pressure compensation opening to the air State indication of the disposal system: The float falls below the "MIN" mark on the sight glass when the disposal system does not work or the pump rate is lower than 25 L/min (Low-flow) or 75 L/min (high-flow).

Connector of the disposal system: ISO 9170-2

### Materials

All materials in contact with exhaled patient gases are autoclavable, except flow sensors (being not capable of being autoclaved), O<sub>2</sub> sensor, and mechanical pressure gauge.

All materials in contact with patient gas are latex free.

### Suction Device

#### Venturi Suction Regulator

Gas source	Air, from system gas source
Minimum flow	20 L/min
Maximum vacuum	≥72 kPa at supply gas pressure of 280 kPa; ≥73 kPa at supply gas pressure of 600 kPa

#### Continuous Suction Regulator

Supply	Negative Pressure Suction
Maximum vacuum	517.5 mmHg to 540 mmHg (69 kPa to 72 kPa) with external vacuum applied of 540 mmHg and 40 L/min free flow
Maximum flow	39 L/min to 40 L/min with external vacuum applied of 540mmHg and 40 L/min free flow
Minimum flow	20 L/min

-----  
Please contact your local Mindray sales representative for the most current information.

Mindray Building, Keji 12th Road South,  
High-tech Industrial Park, Nanshan, Shenzhen 518057, P.R. China  
Tel: +86 755 8188 8998 Fax: +86 755 26582680  
E-mail: intl-market@mindray.com www.mindray.com

**mindray** | Healthcare with reach are registered trademarks or trademarks owned by Shenzhen Mindray Bio-medical Electronics Co., LTD  
© 2017 Shenzhen Mindray Bio-medical Electronics Co., Ltd. All rights reserved. Specifications subject to changes without prior notice.  
P/N: ENG-WATO EX-65 datasheet-210285X4P-20180122

**mindray**